



# **ADPG Pyrophosphorylase Microplate Assay Kit User Manual**

**Catalog # FTA0125**

(Version 1.1A)

Detection and Quantification of ADPG Pyrophosphorylase Activity in  
Tissue extracts, Cell lysate, Cell culture media and Other biological  
fluids Samples.

**For research use only. Not for diagnostic or therapeutic procedures.**

I. INTRODUCTION.....	2
II. KIT COMPONENTS.....	3
III. MATERIALS REQUIRED BUT NOT PROVIDED.....	3
IV. SAMPLE PREPARATION.....	4
V. ASSAY PROCEDURE.....	5
VI. CALCULATION.....	6
VII. TECHNICAL SUPPORT.....	7
VIII. NOTES.....	7

## I. INTRODUCTION

In enzymology, a glucose-1-phosphate adenylyltransferase (EC 2.7.7.27) is an enzyme that catalyzes the chemical reaction

ATP + alpha-D-glucose 1-phosphate → diphosphate + ADP-glucose

Thus, the two substrates of this enzyme are ATP and alpha-D-glucose 1-phosphate, whereas its two products are diphosphate and ADP-glucose.

This enzyme belongs to the family of transferases, specifically those transferring phosphorus-containing nucleotide groups (nucleotidyltransferases).

## II. KIT COMPONENTS

Component	Volume	Storage
96-Well Microplate	1 plate	
Assay Buffer	30 ml x 4	4 °C
Diluent	20 ml x 1	4 °C
Enzyme	Powder x 1	-20 °C
Substrate	Powder x 1	-20 °C
Technical Manual	1 Manual	

**Note:**

**Enzyme:** add 10 ml diluent to dissolve before use.

**Substrate:** add 10 ml diluent to dissolve before use.

## III. MATERIALS REQUIRED BUT NOT PROVIDED

1. Microplate reader to read absorbance at 340 nm
2. Distilled water
3. Pipettor, multi-channel pipettor
4. Pipette tips
5. Mortar
6. Centrifuge
7. Timer
8. Ice

#### **IV. SAMPLE PREPARATION**

##### **1. For tissue samples**

Weigh out 0.1 g tissue, homogenize with 1 ml Assay buffer on ice, centrifuged at 10000g 4 °C for 10 minutes, take the supernatant into a new centrifuge tube and keep it on ice for detection.

## V. ASSAY PROCEDURE

Add following reagents into the centrifuge tube:

Reagent	Sample
Sample	50 $\mu$ l
Substrate	100 $\mu$ l
Mix, incubate at 30°C for 30 minutes, put it into boiling water for 2 minutes. Then keep it on ice for cold. Centrifuged at 10000g 4 °C for 10 minutes, add the supernatant into the microplate.	
Supernatant	100 $\mu$ l
Enzyme	100 $\mu$ l
Mix, measured at 340 nm and record the absorbance of 10th second and 130th second.	

### Note:

1) For unknown samples, we recommend doing a pilot experiment & testing several doses to ensure the readings are within the standard curve range. If the enzyme activity is lower, please add more sample into the reaction system; or increase the reaction time; if the enzyme activity is higher, please dilute the sample, or decrease the reaction time.

## VI. CALCULATION

**Unit Definition:** One Unit of ADPG Pyrophosphorylase activity is defined as the enzyme produces 1 nmol NADPH per minute.

1. According to the protein concentration of sample

$$\begin{aligned} \text{AGPase (U/mg)} &= (\text{OD}_{\text{Sample(130S)}} - \text{OD}_{\text{Sample(10S)}}) / (\epsilon \times d) \times V_{\text{Total}} \times 10^9 / (V_{\text{Sample}} \times C_{\text{Protein}}) \\ &\quad / T1 / T2 \times 1.5 \\ &= 26.8 \times (\text{OD}_{\text{Sample(130S)}} - \text{OD}_{\text{Sample(10S)}}) / C_{\text{Protein}} \end{aligned}$$

2. According to the weight of sample

$$\begin{aligned} \text{AGPase (U/g)} &= (\text{OD}_{\text{Sample(130S)}} - \text{OD}_{\text{Sample(10S)}}) / (\epsilon \times d) \times V_{\text{Total}} \times 10^9 / (W \times V_{\text{Sample}} / \\ &\quad V_{\text{Assay}}) / T1 / T2 \times 1.5 \\ &= 26.8 \times (\text{OD}_{\text{Sample(130S)}} - \text{OD}_{\text{Sample(10S)}}) / W \end{aligned}$$

3. According to the quantity of cells or bacteria

$$\begin{aligned} \text{AGPase (U/10}^4\text{)} &= (\text{OD}_{\text{Sample(130S)}} - \text{OD}_{\text{Sample(10S)}}) / (\epsilon \times d) \times V_{\text{Total}} \times 10^9 / (N \times V_{\text{Sample}} / \\ &\quad V_{\text{Assay}}) / T1 / T2 \times 1.5 \\ &= 26.8 \times (\text{OD}_{\text{Sample(130S)}} - \text{OD}_{\text{Sample(10S)}}) / N \end{aligned}$$

$\epsilon$ : molar extinction coefficient,  $6.22 \times 10^3$  L/mol/cm;

d: the optical path of 96-Well microplate, 0.6 cm;

$C_{\text{Protein}}$ : the protein concentration, mg/ml;

W: the weight of sample, g;

N: the quantity of cell or bacteria,  $N \times 10^4$ ;

$V_{\text{Total}}$ : the total volume of the enzymatic reaction, 0.2 ml;

$V_{\text{Sample}}$ : the volume of sample, 0.05 ml;

$V_{\text{Assay}}$ : the volume of Assay buffer, 1 ml;

T1: the reaction time, 30 minutes.

T2: the reaction time, 2 minutes.

## **VII. TECHNICAL SUPPORT**

For troubleshooting, information or assistance, please go online to [www.cohesionbio.com](http://www.cohesionbio.com) or contact us at [techsupport@cohesionbio.com](mailto:techsupport@cohesionbio.com)

## **VIII. NOTES**